## K063341

#### 510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR § 807.93

Submitter

Implant Innovations, Inc.

4555 Riverside

Palm Beach Gardens, FL 33410

Contact

Jacquelyn A. Hughes, RAC

Director, Regulatory Affairs & Quality Assurance

Implant Innovations, Inc.

4555 Riverside

Palm Beach Gardens, FL 33410

Tel. 561-776-6819 Fax. 561-514 6316

Email jhughes@3implant.com

Date Prepared

October 27, 2006

**Device Name** 

3i OSSEOTITE® Certain® Dental Implants

**Classification Name** 

**Endosseous Dental Implants** 

Device

Class II

Classification

Dental Devices Panel

21 CFR § 872.3640

Predicate

3i OSSEOTITE Certain Dental Implants

Devices

K874590, K935544, K972444, K980549, K983347,

K022009, K031475, K030164, K031362, K041402,

K051189, K051461

Performance

Performance standards have not been established by the

FDA under Section 514 of the Federal Food, Drug and

Cosmetic Act.

### **Device Description**

The 3i OSSEOTITE® Certain® Dental Implants are provided with the proprietary OSSEOTITE acid-etched surface which has been in commercial distribution since market clearance in 1995. Implants are offered in tapered and parallel-walled/straight designs, each design providing offerings for internal hex connections. Additionally, the implants are offered in an expanded platform, lateralized design with an internal connection Implants are offered in diameters of 3.25, 3.75, 4.0, 5.0, and 6.0 in varying lengths from 7 mm to 20 mm. Size appropriate cover screws are offered with each implant.

#### Indications for Use

3i dental implants are intended for surgical placement in the upper or lower jaw to provide a means for prosthetic attachment in single tooth restorations and in partially or fully edentulous spans with multiple single teeth, or as a terminal or intermediary abutment for fixed or removable bridgework, and to retain overdentures.

In addition, when a minimum of 4 implants,  $\geq$  10mm in length, are placed in the mandible and splinted in the anterior region, immediate loading is indicated.

### Technological Characteristics

The design features and functions are identical to the currently available OSSEOTITE Certain, OSSEOTITE NT® Certain, OSSEOTITE XP® Certain and Certain® PREVAIL<sup>TM</sup>, and *3i* Implant Innovations implants and cover screws.

#### **Performance Testing**

Laboratory testing was conducted to determine device functionality and conformance to design input requirements, as well as FDA's Class II special controls guidance document: Root-form Endosseous Dental Implants and Endosseous Dental Abutments. Risk analysis was conducted in accordance with ISO 14971. Specific performance claims for enhanced performance in poor bone (K980549), early loading (K983347), and use in smokers (K022009), and immediate loading (K030164) have received market clearance based on the OSSEOTITE clinical data in the respective submissions.

Implant Innovations, Inc. 510(k) Premarket Notification -3i OSSEOTITE® Certain ®Dental Implants

#### Conclusion

The 3i OSSEOTITE Certain Dental Implants are substantially equivalent to the dental implants described in the premarket notification submissions for the predicate devices K879450, K935544, K972444, K980549, K983347 K022009, K031475, K030164, K031362, K041402, K051189, K051461.





Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

APR 3 0 2007

Ms. Jacquelyn A. Hughes Director, Regulatory Affairs & Quality Assurance Implant Innovations, Incorporated 4555 Riverside Drive Palm Beach Gardens, Florida 33410

Re: K063341

Trade/Device Name: 3i OSSEOTITE® Certain® Dental Implants

Regulation Number: 872.3640

Regulation Name: Endosseous Dental Implant

Regulatory Class: II Product Code: DZE Dated: April 12, 2007 Received: April 13, 2007

#### Dear Ms. Hughes:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <a href="http://www.fda.gov/cdrh/industry/support/index.html">http://www.fda.gov/cdrh/industry/support/index.html</a>.

Sincerely yours,

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital, Infection Control and Dental Devices Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

# **Indications for Use**

510(k) Number (if known): <u>K063341</u>
Device Name: 3i OSSEOTITE® Certain® Dental Implants
Indications for Use:
3i dental implants are intended for surgical placement in the upper or lower jaw to provide a means for prosthetic attachment in single tooth restorations and in partially or fully edentulous spans with multiple single teeth, or as a terminal or intermediary abutment for fixed or removable bridgework, and to retain overdentures.
In addition, when a minimum of 4 implants, $\geq$ 10mm in length, are placed in the mandible and splinted in the anterior region, immediate loading is indicated.
Prescription Use X (Part 21 CFR 801 Subpart D)  AND/OR Over-The-Counter Use (21 CFR 801 Subpart C)
(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)
Concurrence of CDRH, Office of Device Evaluation (ODE)
Susan Punny
i elicin of Anesthesiology, General Hospital, iadodion Control, Dental Devices
C10(k) Number: 166384 Page of